

LABORATORY SAFETY INSPECTION CHECKLIST

Date of Inspection 12/5/12

Section Inspected: AMHERST DRUGS

Inspector(s): REBECCA PONTES.

Physical Inspection

Comments

- ☒ Unobstructed exits
- ☒ All storage 18" below ceiling
- ☒ Unobstructed aisles
- ☒ Workbenches clean & orderly
- Chemical/Extraction hoods**
 - ☒ Free from clutter/obstruction
 - ☒ Chemicals placed 6 inches away from sash and 3 inches from back wall

Chemical Inspection

- ☒ Bottles clearly & properly labeled
- ☒ Chemicals stored properly
- ☒ Waste bottles clearly labeled and stored

General Safety

- ☒ Fire extinguisher accessible and inspected
Last inspection date: 6/2012
- ☒ Eyewash stations unobstructed and inspected weekly. Last inspection date: Room 236 12/3/12
Room 245 12/4/12
- ☒ Safety shower unobstructed and inspected quarterly. Last inspection date: Room 236 12/4/12
and 245
- ☐ AED Last inspection date: N/A
- ☒ First aid kit stocked
- ☒ Sinks unobstructed and clear
- ☒ Spill kit accessible and stocked

Personal Protective Equipment

- ☒ Lab coats used in laboratory working areas
- ☒ Appropriate eyewear protection used
- ☒ Appropriate glove protection used
- ☒ Evacuation plan posted in plain view
- ☒ MSDS available in unit
- ☒ Safety manual is accessible
- ☒ Biohazard bins not overfilled
- ☒ Cleaning materials are accessible, unexpired
- ☒ No obvious food or drink being consumed in laboratory area